

QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM

CSF Neurotransmitters

ERNDIM Workshop, 21st & 22nd November 2017, Manchester, UK

Overview



- Short explanation of the scheme
- Review of last years samples
- Discussion questions, concerns, suggestions

• This is a workshop so please feel free to interupt, ask questions and participate

CSF Scheme



- Started in 2014
- The purpose was to set up an EQA scheme so labs measuring CSF Homovanillic acid, 5-hydroxyindole acetic acid, 3-methyl dopa and 5-hydroxytryptophan could compare their results against other labs. This helps for accreditation purposes but also give greater confidence in results given to patients.
- Interpretation of metabolite patterns was added as considered as important as quantitation
- Currently ~30 labs participating a mix of HPLC-ECD and LC-MS methods
- Simon Heales and I are scientific advisors

The samples



- 8 lyophilised samples a year 4 samples in duplicate
- Samples a mix of pooled CSF and artificial
- Samples prepared in London and lyophilised in the Netherlands
- Tested for stability before sending to participants

Scoring



- Scoring based on 4 criteria accuracy, precision, linearity and recovery for each metabolite
- Scoring done automatically by website
- Scores outside 3 SDs are flagged, as are no results
- A score of >25 (25% of results flagged) will get a warning letter
- More details on the ERNDIM website and in Cas's talk tomorrow

Scoring example

ERNDIMQA - ANNUAL REPORT ERNDIM Quant.org.acids 2014 ERN0332 - DMC University Laboratories - Biochemical Genetics Laboratory Methodset : GC-MS Recovery Data All Labs Accuracy Precision Linearity (mean) (CV% duplicates) (r) (%added analyte) Analyte All Labs Yourlah Interlab CV Yourlab Your Lab Allabs All abs Your Lab Allabs 11 OH Glutaric acid 152 32 0% 20.0% 0.077 0.900 09% 93% 77 35,0% 70.4 28.3% 15 0% 0 9 14 0.985 117% 101% methylglutaric acid MP 77 23 8% 3-OI I-3 methylglutaric acid TR 00.9 FR 24.7% ER. 0.903 FR 77% 65 62.4% 3-OH-Isobulyric acid OR 130 OR 20.4% OR 0.979 OR 71% 53 37.0% **FR** 70.1 FR 39.3% **FR** 0.940 FR 112% 73 55.5% 3-OI I-Isovaleric acid 4-OH-Butyric acid FR 72.8 FR 31 2% FR 0.982 FR 70% 63 56.7% MP 178 31.6% 12.6% 0.050 0.003 85% 101% 85 26,4% dip c acit Creatinine 3258 3.4% 0.000 0% 67 5.25% FR 278 64 D.L. Glycenic acid FR 22.0% FR 0.985 FR 83% 47 0% ORER 50.0 ORFR 17.1% ORER 0.992 ORER 99% 89 36.3% Ethymalonic acid MP 60.0 19.5% 14 5% 0.991 0.991 75% 100% 78 28 5% umaric aci Glutaric acid MP 112 44.0% 13.1% 0.994 86% 101% 90 21,8% OR. OR OF OR 77 0.953 Glycolic acid 236 20.6% 89% 38 0% MP 17.0 29.3% 29.9% 0.949 0.950 111% 100% 64 37.2% 198 0.992 103% 73 40.1% Keto-alutaric acid 21.4% 117 19 8% 0.995 81% 53 47 4% Malic acid 330 othylmplonic o 305 31.2% 11 4% 0.005 0.995 9556 102% 104 27.5% Mevalonic acid FR 132 0.992 FR 82% 56 46 1% FR 217% FR 353 322 0.879 0.950 85% 73 44.2% 61.6% 23.0% 108% proclutamic acie MP 55.4 15.2% 19.4% 0.987 0.957 91% 101% 83 35,9% Suberic acid 04 32 4% 172 11.5% 16.3% 0.992 0.990 105% MP 57 33.6 17.0% 24,5% 0.975 0.907 54% 05% 46 0% Vanilactic acid 30.5 19.4% 0.975 90% 30 33.0% 270 80% 71 Overal 260 20.0% 20.1% 0.044 0.040 88% 37.7%



QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM

Calculation Score

Total Number Relevant Boxes

-23 Analytes

-4 analytes no results (empty boxes)
-5 analytes Few Results (FR)
Relevant Analytes: 23 - (4+5) = 14
Relevant Boxes: 14 X 4 = 56

Total Number (weighed) Flags

- 4 Yellow boxes = 4/4 = 1
- 8 Purple boxes = 8/4 = 2
- 8 Amber boxes = 8/1 = 8
- 3 Red boxes = 3/1 = 3

Total weighed flags: 1+2+8+3 = 14

Score =	Total number flags	= 14 x 100% = 25
	Total number boxes	56


www.erndim.org

MP - missing pair; OR - outlying result; ORFR - outlying/few results; FR - few results (<6/8); Red flag - poor performance

This year's samples 1 and 8



QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM




3 or 4 year old Normal ratio and metabolites within reference ranges

Samples 1 and 8




QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM



Samples 2 and 5



QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM




5 or 7 year old Normal ratio but elevated 3-methyl dopa and 5-hydroxtryptophan

Samples 2 and 5




QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM



Samples 3 and 7



QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM




5 or 7 year old Normal ratio All metabolites within reference ranges

Samples 3 and 7




QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM



Samples 4 and 6



QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM




3 or 4 year old Normal ratio Elevated 3-methyl dopa

Samples 4 and 6



QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM



Overview of performance



- Results for 5HIAA, HVA and ratio generally good with CVs of ~10% between labs
- Much greater variation for 3-methyl dopa and 5-hydroxytryptophan >30%

Methodset :	133% + hpic ecd	International Contraction	The Manual Ar							
Accuracy (mean)			Precision (CV% duplicates)		Linearity (r)		Recovery (%added analyte)		Data All Labs	
Analyte	Your Lab	All Labs	Your Lab	All Labs	Your Lab	All Labs	Your Lab	All Labs	n	Interlab CV
3-methyl dopa (3-MD)	372	369	8.0%	11.3%	0.998	0.995	101%	100%	24	34.0%
5-OH-Indolacetic acid (5- HIAA)	138	145	5.3%	9.9%	0.989	0.932	104%	104%	27	13.2%
5-OH-Tryptophan (5-HTTP)	57.8	49.0	10.5%	18.4%	0.999	0.996	115%	100%	19	66.1%
Homovanillic acid (HVA)	336	317	2.2%	6.7%	0.989	0.985	106%	101%	27	11.2%
HVA:5HIAA ratio	2.38	2.28	4.5%	8.3%	0.959	0.914	95%	103%	23	12.6%
Overall	181	176	6.1%	10.9%	0.987	0.964	104%	102%	24	27.4%

Possible reasons for poor performance



QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM

Random mistakes

- Reconstitution too much, too little, not long enough
- Wrong sample analysed


Systematic

- Poor control of assay standards, QC or voltamagram problems
- Mis-identification of metabolites variation in pH can shift peaks
- Co-elution of peaks

Voltammagram



QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM




mV

Very important to check voltammagram and peak areas regularly. 3-methyl dopa hardest to oxidise so changes to voltammagram most likely to affect 3-methyl dopa first. This may explain greater variation for 3-methyl dopa.

Mis-identification



QUALITY ASSURANCE IN LABORATORY TESTING FOR IEM



Other peaks in CSF could be mis-identified

3MD and 5HTP if pH increased

3MD and 5HTP if pH lowered

Any suggestions?



Suggestions, complaints, improvements, discussion
Possible changes to the scheme - greater emphasis on

interpretation?