

ANNUAL REPORT 2023

Scheme Organiser	Scientific Advisor	Website for reporting results	Administration office
<p>Dr. C.W. Weykamp Streekziekenhuis Koningin Beatrix MCA Laboratory Beatrixpark 1 7101 BN Winterswijk The Netherlands e-mail: mca.office@skbwinterswijk.nl</p>	<p>Dr. Rachel Carling Biochemical Sciences, Viapath 4th Floor, North Wing St Thomas' Hospital London SE1 7EH United Kingdom e-mail: Rachel.Carling@viapath.co.uk</p> <p>Prof. Stuart Moat Medical Biochemistry, Immunology & Toxicology University Hospital of Wales Heath Park Cardiff CF14 4XW United Kingdom e-mail: Stuart.Moat@wales.nhs.uk</p>	<p>Mrs. Irene de Graaf Streekziekenhuis Koningin Beatrix MCA Laboratory Beatrixpark 1 7101 BN Winterswijk The Netherlands e-mail: i.degraaf@skbwinterswijk.nl</p>	<p>ERNDIM Administration Office c/o EMQN CIC, Unit 4, Enterprise House, Manchester Science Park Pencroft Way Manchester M15 6SE United Kingdom. e-mail: admin@erndim.org</p>

Published: London-Winterswijk, 9th January 2024¹

1. **Purpose**

The purpose of the ERNDIM External Quality Assurance Scheme for Special Assays in dried blood spots is the monitoring of the analytical quality of the quantitative assay of a range of analytes in dried blood spots in laboratories involved in the diagnosis of patients with inherited metabolic disorders. For details see www.erndim.org / www.ERNDIMQA.nl

2. **Participants**

A total of 112 datasets have been submitted. One laboratory did not submit results at all.

3. **Design**

The Scheme has been designed, planned and co-ordinated by the scientific advisors Dr. Rachel Carling and Professor Stuart Moat and Dr. C.W. Weykamp as scheme organizer (on behalf of the MCA Laboratory), all three appointed by and according to the procedure of the ERNDIM Board. The design includes samples and reports to provide information with a balance between short-term and long-term reports and between detailed and aggregated information. As a subcontractor of ERNDIM, the MCA Laboratory prepares and dispatches EQA samples to the scheme participants and provide a website for on-line submission of results and access to scheme reports.

Samples

The scheme consisted of eight dried blood spots, comprising four identical pairs. In addition, a standard material was included for analysis with each distribution. All samples were prepared from the same basic whole blood but with varying amounts of

¹ If this Annual Report is not Version 1 for this scheme year, go to APPENDIX 1 for details of the changes made since the last version of this document

added analytes. Blood was depleted to create low concentrations but not to a zero level. Thus concentrations of the samples is the remaining physiological concentration plus the spiked amount. The analytes and their source, as well as the added amounts, are shown in the table below for each pair. Samples have been tested for stability and homogeneity according to ISO 13528.

Table 1.

Analyte	Source:	Added Amounts ($\mu\text{mol/L}$)				
		Sample Pair 2023. 01 & 13	Sample Pair 2023. 03 & 09	Sample Pair 2023. 05 & 11	Sample Pair 2023. 07 & 15	Extra standard sample
Allo-isoleucine	Sigma I8454	4,9	10,1	25,0	50,0	10,0
Free Carnitine	Sigma C0283	3,0	18,0	38,0	78,0	23,0
Total Homocysteine	Sigma H6010	170,1	111,1	51,6	21,9	21,8
Isoleucine	Roth 3922.1	783,7	384,2	83,9	33,5	484,1
Leucine	Roth 3984.1	0,0	48,8	350,7	750,1	452,8
Methionine	Sigma 64319	739,9	490,6	239,9	39,9	89,8
NTBC	Sigma PHR1731	0,0	10,0	20,0	50,0	25,0
Phenylalanine	Sigma 78019	1975,9	975,8	477,0	75,1	575,7
Succinylacetone	Sigma D1415	20,0	0,0	10,0	5,0	15,0
Tyrosine	Sigma 93829	728,5	353,2	128,0	28,7	579,5
Valine	Roth 4879.1	447,3	194,6	71,7	20,5	70,8

Reports

All data-transfer, the submission of data as well as request and viewing of reports, proceeded via the interactive website www.erndimqa.nl which can also be reached through the ERNDIM website (www.erndim.org). The results of your laboratory are confidential and only accessible to you (with your name and password). The anonymised mean results of all labs are accessible to all participants. Statistics of the respective reports are explained in the general information section of the website.

An important characteristic of the website is that it supplies short-term and long-term reports.

Short-term reports on the ten individual specimens are available two weeks after the submission deadline and provide up-to-date information on analytical performance. Although it is technically possible to produce reports immediately there is a delay of 14 days to enable the scientific advisor to inspect the results and add comments to the report when appropriate.

The **annual long-term report** is based on the design-anchored connection between samples which enables a range of analytical parameters (accuracy, precision, linearity, recovery and inter-lab dispersion) to be reported once the annual cycle has been completed. The annual report is discussed below.

A second important characteristic of the website is the wide range in aggregation of results which permits labs to make an individual choice for detailed and/or aggregated reports. The most detailed report which can be requested from the website is the "Analyte in Detail" which shows results of a specific analyte in a specific sample (176 such Analyte-in-Detail-reports can be requested in the 2023 cycle). A more condensed report is the "Cycle Review" which summarizes the performance of all analytes in a specific sample (16 such Cycle-Review-Reports can be requested in 2023). The most comprehensive report is the Annual Report which summarizes the performance of all analytes in each of the four pairs of samples. 1 such Annual-Report can be requested in 2023.

4. Discussion of Results in the Annual Report 2023

In this section the results of the annual report 2023 are summarised in terms of accuracy, recovery, precision, linearity, inter-laboratory co-efficient of variation (CV) and relations between these parameters. Please keep at hand your annual report from the Interactive Website when you read the “guided tour” below and keep in mind that we only discuss the results of “all labs”: it is up to you to inspect and interpret the specific results of your laboratory.

4.1 Accuracy

A first approach to describe accuracy is to compare the mean outcome of the eight samples from your lab with the mean outcome from all labs. This is done in the first column of the annual report. It can be seen that the mean laboratory mean concentration for alloisoleucine is 21.9 µmol/L.

4.2 Recovery

A second approach to describe accuracy is the percentage recovery of added analyte. In this approach, recovery is measured relative to the weighed amount of analyte used to enrich the sample. The correlation between the weighed quantities added to the samples (on the x-axis) and your measured quantities (on the y-axis) has been calculated. The slope of the correlation multiplied by 100% is your recovery of the added amounts. Outcome for your laboratory in comparison to median outcome of all laboratories is shown in the column “Recovery” in the Annual Report. The all laboratory mean recovery ranges from 31% for succinylacetone to 107% for total homocysteine.

4.3 Precision

Reproducibility is an important parameter for quality in the laboratory and is encountered in the schemes’ design. Samples are prepared in pairs and each pair can be regarded as duplicates from which CV’s can be calculated (intra-laboratory CV as an indicator of reproducibility). Outcome for your lab in comparison to the median of all labs is shown in column “Precision” of the Annual Report. The all laboratory mean imprecision ranges from 6.7% for tyrosine to 16.7% for NTBC.

4.4 Linearity

Linearity over the analytical range is another important parameter for analytical quality. Again this is encountered in the Schemes’ design. With weighed quantities on the x-axis and your measured quantities on the y-axis the coefficient of regression (r) has been calculated. Outcome for your lab in comparison to the median of all labs is in the column “Linearity” of the annual report. It can be seen that the coefficient of regression for phenylalanine is 0.991.

4.5 Inter-laboratory CV

For comparison of outcome for one patient in different hospitals and for use of shared reference values it is important to have a high degree of inter-laboratory harmonization. Part of the schemes’ design is to monitor this by calculating the inter-laboratory CV. This, along with the number of laboratories who submitted results, is shown in the column “Data all Labs” in the Annual Report. It can be seen that most laboratories submitted results for phenylalanine (n = 93) whereas only fourteen labs submitted results for NBTC. The mean inter-laboratory CV ranges from 14.5% for leucine to 67.3% for succinylacetone.

4.6 **Cross Sectional Relations**

The various parameters as described above often have an inter-relation: more than one parameter directs towards good or poor analytical performance.

4.7 **Your performance: Flags**

In order to easily judge performance of individual laboratories the annual report of an individual laboratory may include flags in case of poor performance for accuracy, precision, linearity and recovery. Analytes with satisfactory performance for at least three of the four parameters (thus no or only one flag) receive a green flag. Thus a green flag indicates satisfactory performance for analysis of that particular analyte. Criteria for flags can be found in the general information on the website (on this website under general information; interactive website, explanation annual report).

4.8 **Poor Performance Policy**

It is evident that there is considerable variation in the overall performance of individual laboratories. Table 2 shows the percentage of flags observed. 47% of the laboratories have no flag at all and thus have attained excellent overall performance. In contrast, at the other extreme, 5% of laboratories have more than 25% red flags. However, it should be noted that not all laboratories return results for all analytes. Following intensive discussion within the ERNDIM board and Scientific Advisory Board (SAB) and taking into account feedback from participants we have been able to agree on a harmonised scoring system for the various branches of the Diagnostic Proficiency schemes and qualitative schemes. This has now been in place for more than 10 years. In parallel to this the SAB has agreed levels of adequate performance for all the schemes and these will be re-evaluated annually. The scoring systems have been carefully evaluated by members of the SAB and have been applied to assess performance in our schemes from 2007 onwards. The ERNDIM Board has decided that the Scientific Advisor will judge the performance of the individual laboratories based on these levels of satisfactory performance and issue a letter of advice of failure to achieve satisfactory performance to those laboratories which do not achieve satisfactory performance. The letter is intended to instigate dialogue between the EQA scheme organiser and the participating laboratory in order to solve any particular analytical problems in order to improve quality of performance of labs in the pursuit of our overall aim to improve quality of diagnostic services in this field.

If your laboratory is assigned poor performance and you wish to appeal against this classification please email the ERNDIM Administration Office (admin@erndim.org), with full details of the reason for your appeal, within one month receiving your Performance Support Letter. Details of how to appeal poor performance are included in the Performance Support Letter sent to poor performing laboratories.

Table 2. Percentage Flags

% Red Flags seen in Annual Report	Percentage Labs In this Category	Cumulative Percentage Of Labs
>25%	5%	5%
25%	5%	10%
20 – 25%	1%	11%
15 – 20%	5%	16%
10 – 15%	11%	27%
5 – 10%	19%	46%
0 – 5%	7%	53%
0%	47%	100%

4.9 Certificates

As for other schemes the performance as it is indicated by the red/green flags in the individual laboratories annual report is summarised in the annual participation certificate. The certificate lists the total number of special assays in the scheme, the number for which results have been submitted and the number for which satisfactory performance has been achieved. It is important to bear in mind that the certificate has to be backed up by the individual annual report in the case of internal or external auditing.

4.10 Additional Specific Remarks of the Scientific Advisor

The scheme results, which are consistent with those seen in previous years, highlight the poor performance of NTBC, total homocysteine and succinyl acetone. Although only a relatively small number of laboratories return results on these analytes, n=14, 33 and 35 respectively, there is evidently a need to improve performance of these analytes.

5. Summary

The Annual Report deals with analytical performance in terms of accuracy, precision, linearity, recovery and inter-lab CV. All parameters (intra-laboratory CV, linearity, recovery, inter-laboratory CV and the number of participating laboratories) are broadly comparable to that seen in 2022. Intra-laboratory variation was acceptable for all analytes; mean variance 102%, range 6.3 – 16.7%. Inter-laboratory imprecision highlights the ongoing challenges associated with the measurement of some of these analytes in dried blood spots; mean inter-laboratory variation 25.1%, ranges from 14.5% for leucine, to 67.3% for succinylacetone.


6. Preview Scheme 2024

The design of the 2024 scheme is essentially the same as in 2023.

7. Questions, Comments and Suggestions

If you have any questions, comments or suggestions please address to the scientific advisor of the scheme Dr. Rachel Carling (Rachel.Carling@viapath.co.uk) and/or to the scheme organiser Dr. C. W. Weykamp (mca.office@skbwinterswijk.nl)

London, 09/01/2024



Dr. Rachel Carling
Scientific Advisor

Please note:

This annual report is intended for participants of the ERNDIM Special Assays in dried blood spots scheme. The contents should not be used for any publication without permission of the scheme advisor.

The fact that your laboratory participates in ERNDIM schemes is not confidential. However, the raw data and performance scores are confidential and will be shared within ERNDIM for the purpose of evaluating your laboratory performance, unless ERNDIM is required to disclose performance data by a relevant government agency. For details, please see the terms and conditions in the ERNDIM Privacy Policy on www.erndim.org.

APPENDIX 1. Change log (changes since the last version)

Version Number	Published	Amendments
1	9 th January 2024	<ul style="list-style-type: none">• 2023 annual report published

END